



CERTIFICATE OF ACCREDITATION

ANSI National Accreditation Board

11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

Morgan Schaffer Ltd.

**8300 Saint-Patrick Street, Suite 150
LaSalle, QC H8N 2H1**

has been assessed by ANAB and meets the requirements of international standard

ISO 17034:2016

while demonstrating technical competence in the field of

REFERENCE MATERIAL PRODUCER

Refer to the accompanying Scope of Accreditation for information regarding the types of materials to which this accreditation applies

AR-2126

Certificate Number



ANAB Approval

Certificate Valid Through: 03/08/2021
Version No. 004 Issued: 01/18/2019





SCOPE OF ACCREDITATION TO ISO 17034:2016

Morgan Schaffer Ltd

8300 Saint-Patrick Street, Suite 150

LaSalle, QC H8N 2H1

Denis Lafrance 514-739-1967

dlafrance@morganschaffer.com

www.morganschaffer.com

REFERENCE MATERIAL PRODUCER

Valid to: **March 8, 2021**

Certificate Number: **AR-2126**

Chemical

| Sub-Category of Reference Material | ILAC RM Category | Class or Type of Reference Materials Produced (Include Range Where Applicable) | Methods or Techniques Used in the RMP Laboratory (if Appropriate) |
|---|------------------|---|---|
| Gases (Certified Reference Material) | A7.1 | Gas mixture in dielectric fluids (5, 10, 50, 100, 500 and 1000 ppm ($\mu\text{g/g}$) of H_2 , CH_4 , C_2H_6 , C_2H_4 , C_2H_2 , CO_2 , O_2 , N_2 , and CO) | Gas Chromatography |
| Moisture (Certified Reference Material) | A3.5 | Moisture in dielectric fluids 15 and 30 ppm ($\mu\text{g/g}$) | Karl Fisher Titrator |

Notes:

1. Please contact the RMP organization for more information on CRM uncertainty values, Ucrm values, and other specific lot values. Some of this information may also be available on the RMP's website.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AR-2126.



Vice President

