



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Northeast Metrology Corp.
4490 Broadway
Depew, NY 14043

Fulfills the requirements of

ISO/IEC 17025:2017

and national standard

ANSI/NCSL Z540-1-1994 (R2002)

In the fields of

CALIBRATION and TESTING

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to be 'J. Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 06 June 2028

Certificate Number: ACT-1116



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

Northeast Metrology Corp.

4490 Broadway

Depew, NY 14043

Mike Moczerniak 716-827-3770 mike@nemcal.com

CALIBRATION AND TESTING

ISO/IEC 17025 Accreditation Granted: **20 May 2026**

Certificate Number: **ACT-1116**

Certificate Expiry Date: **06 June 2028**

CALIBRATION

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source ^{1,2}	Up to 330 mV (0.33 to 3.3) V (3.3 to 33) V (33 to 330) V (330 to 1 020) V	7.1 μ V 40 μ V 0.4 mV 5.9 mV 20 mV	Comparison to Multiproduct Calibrator
DC Voltage – Measure ^{1,2}	Up to 200 mV (0.2 to 2) V (2 to 20) V (20 to 200) V (200 to 1 020) V	7.6 μ V 37 μ V 0.32 mV 7 mV 55 mV	Comparison to 8.5 Digit Multimeter
DC Current – Source ^{1,2}	Up to 330 μ A (0.33 to 3.3) mA (3.3 to 33) mA (33 to 330) mA (0.33 to 1.1) A (1.1 to 11) A (11 to 20.5) A	72 nA 0.38 μ A 3.5 μ A 35 μ A 0.61 mA 5.1 mA 24 mA	Comparison to Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure ^{1,2}	Up to 200 μ A (0.2 to 2) mA (2 to 20) mA (20 to 200) mA (0.2 to 2) A	0.1 μ A 0.99 μ A 10 μ A 0.13 mA 2.2 mA	Comparison to 8.5 Digit Multimeter
DC Current – Measure ^{1,2}	(2 to 10) A	16 mA	Comparison to 6.5 Digit Multimeter
DC Current – Measure ^{1,2}	(400 to 2 000) A	19 A	Comparison to Clamp-on Meter
AC Voltage – Source ^{1,2}	Up to 33 mV 10 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (33 to 330) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (0.33 to 3.3) V (10 to 45 Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (3.3 to 33) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	16 μ V 15 μ V 40 μ V 0.13 mV 0.3 mV 0.13 mV 0.13 mV 73 μ V 0.13 mV 0.28 mV 0.74 mV 1.3 mV 1.3 mV 0.72 mV 1.1 mV 2.3 mV 8.9 mV 9.1 mV 14 mV 9.6 mV 13 mV 32 mV	Comparison to Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ^{1,2}	(33 to 330) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (330 V to 1 020) V 45 Hz to 5 kHz (5 to 10) kHz	0.11 V 73 mV 97 mV 0.11 V 0.74 V 0.26 V 0.32 V	Comparison to Multiproduct Calibrator
AC Voltage – Measure ^{1,2}	Up to 200 mV (20 to 50) Hz (50 to 100) Hz 100 Hz to 10 kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (100 to 200) kHz 200 kHz to 2 MHz (0.22 to 2) V (20 to 100) Hz 100 Hz to 10 kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (100 to 200) kHz 200 kHz to 2 MHz (2 to 20) V (20 to 50) Hz 50 Hz to 1 kHz (1 to 5) kHz (5 to 25) kHz (25 to 50) kHz (50 to 100) kHz (20 to 200) V (20 to 50) Hz 50 Hz to 1 kHz (2 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz	1.6 mV 0.21 mV 99 μV 0.11 mV 0.76 mV 1.9 μV 5.1 mV 13 mV 2.1 mV 1 mV 1.1 mV 7.6 mV 19 mV 51 mV 0.13 V 37 mV 11 mV 13 mV 16 mV 76 mV 0.19 V 0.38 V 0.12 V 0.14 V 0.17 V 0.77 V 1.9 V	Comparison to 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ^{1,2}	(200 to 750) V		Comparison to 8.5 Digit Multimeter
	(20 to 50) Hz	1.5 V	
	50 Hz to 1 kHz	1.1 V	
	(2 to 10) kHz	1.2 V	
	(10 to 25) kHz	1.4 V	
AC Current – Source ^{1,2}	(29 to 330) μ A		Comparison to Multiproduct Calibrator
	(20 to 45) Hz	0.52 μ A	
	45 Hz to 1 kHz	0.52 μ A	
	(1 to 5) kHz	0.87 μ A	
	(5 to 10) kHz	2.6 μ A	
	(10 to 30) kHz	3.9 μ A	
	(0.33 to 3.3) mA		
	(20 to 45) Hz	3.2 μ A	
	45 Hz to 1 kHz	3.2 μ A	
	(1 to 5) kHz	5.2 μ A	
	(5 to 10) kHz	16 μ A	
	(10 to 30) kHz	18 μ A	
	(3.3 to 33) mA		
	(20 to 45) Hz	16 μ A	
	45 Hz to 1 kHz	16 μ A	
	(1 to 5) kHz	28 μ A	
	(5 to 10) kHz	68 μ A	
	(10 to 30) kHz		
	(33 to 330) mA	0.13 mA	
	(10 to 20) Hz	0.63 mA	
	(20 to 45) Hz	0.16 mA	
	45 Hz to 1 kHz	0.16 mA	
	(1 to 5) kHz	0.4 mA	
	(5 to 10) kHz	0.75 mA	
	(10 to 30) kHz	1.5 mA	
	(0.33 to 1.1) A		
	(10 to 45) Hz	0.69 mA	
	45 Hz to 1 kHz	0.69 mA	
	(1 to 5) kHz	7.8 mA	
	(5 to 10) kHz	33 mA	
(1.1 to 3) A			
(10 to 45) Hz	1.9 mA		
45 Hz to 1 kHz	1.9 mA		
(1 to 5) kHz	19 mA		
(5 to 10) kHz	78 mA		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ^{1,2}	(3 to 11) A		Comparison to Multiproduct Calibrator
	(45 to 100) Hz	8.7 mA	
	100 Hz to 1 kHz	12 mA	
	(1 to 5) kHz	0.3 A	
	(11 to 20.5) A		
	(45 to 100) Hz	31 mA	
Resistance – Source ^{1,2} (Simulation)	100 Hz to 1 kHz	26 mA	Comparison to Multiproduct Calibrator
	(1 to 5) kHz	0.62 A	
	Up to 11 Ω	1.7 mΩ	
	(11 to 33) Ω	2.5 mΩ	
	(33 to 110) Ω	4.8 mΩ	
	(110 to 330) Ω	11 mΩ	
	330 Ω to 1.1 kΩ	33 mΩ	
	(1.1 to 3.3) kΩ	0.11 Ω	
	(3.3 to 11) kΩ	0.32 Ω	
	(11 to 33) kΩ	1.1 Ω	
	(33 to 110) kΩ	3.2 Ω	
	(110 to 330) kΩ	12 Ω	
	330 kΩ to 1.1 MΩ	37 Ω	
	(1.1 to 3.3) MΩ	0.51 kΩ	
	(3.3 to 11) MΩ	0.5 kΩ	
(11 to 33) MΩ	0.97 kΩ		
(33 to 110) MΩ	62 kΩ		
(110 to 330) MΩ	21 kΩ		
330 MΩ to 1.1 GΩ	6.4 MΩ		
Resistance – Measure ^{1,2}	Up to 20 Ω	1.8 mΩ	Comparison to 8.5 Digit Multimeter
	(20 to 200) Ω	8.2 mΩ	
	(0.2 to 2) kΩ	41 mΩ	
	(2 to 20) kΩ	0.44 Ω	
	(20 to 200) kΩ	6.5 Ω	
	(0.2 to 2) MΩ	0.11 kΩ	
	(2 to 20) MΩ	4.3 kΩ	
	(20 to 200) MΩ	1.9 MΩ	
(0.2 to 1) GΩ	18 MΩ		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Source ^{1,2} (Simulation)	(0.19 to 3.3) nF (3.3 to 11) nF (11 to 33) nF (110 to 330) nF (0.33 to 1.1) μF (1.1 to 3.3) μF (3.3 to 11) μF (11 to 33) μF (33 to 110) μF (110 to 330) μF (0.33 to 1.1) mF (1.1 to 3.3) mF (3.3 to 11) mF (11 to 33) mF (33 to 110) mF	14 pF 40 pF 0.13 nF 1.1 nF 4 nF 13 nF 40 nF 0.16 μF 0.64 μF 2.5 μF 8.1 μF 25 μF 57 μF 0.51 mF 1.4 mF	Comparison to Multiproduct Calibrator
Electrical Simulation of Thermocouple Indicating Devices ¹	Type E (-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1 000) °C Type J (-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1 200) °C Type K (-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1 000) °C (1 000 to 1 372) °C Type R (0 to 250) °C (250 to 400) °C (400 to 1 000) °C (1 000 to 1 767) °C	0.88 °C 0.74 °C 0.73 °C 0.74 °C 0.75 °C 0.66 °C 0.62 °C 0.62 °C 0.62 °C 0.64 °C 0.68 °C 0.63 °C 0.62 °C 0.65 °C 0.72 °C 3.6 °C 3.6 °C 3.6 °C 3.6 °C	Comparison to Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices ¹	Type S		Comparison to Multiproduct Calibrator
	(0 to 250) °C	3.6 °C	
	(250 to 1 000) °C	3.6 °C	
	(1 000 to 1 400) °C	3.6 °C	
	(1 400 to 1 767) °C	3.6 °C	
	Type T		
	(-250 to -150) °C	0.87 °C	
	(150 to 0) °C	0.65 °C	
(0 to 120) °C	0.62 °C		
(120 to 400) °C	0.62 °C		

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gage Blocks ³	Up to 0.09 in (0.1 to 4) in (> 4 to 20) in	4 μin (2.6 + 1.8L) μin (0.5 + 2.1L) μin	Comparison to Grade 0 & 00 Gage Blocks, Gage Block Comparator
1D Measuring Machine ¹ (lead screw)	Up to 1 in	8.6 μin	Comparison to Grade 0 Gage Blocks, Precision Ball, Oz Testers
Outside Micrometer ^{1,3} (5 μin resolution) (50 μin resolution) (100 μin resolution) (0.001 in resolution)	Up to 1 in Up to 6 in Up to 40 in Up to 40 in	8.8 μin (27 + 1.1L) μin (81 + 1.8L) μin (600 + 0.7L) μin	Comparison to Grade 0 Gage Blocks
Height Gage ^{1,3} (0.000 004 in resolution) (0.000 01 in resolution) (0.000 05 in resolution) (0.000 1 in resolution) (0.000 5 in resolution) (0.001 in resolution)	Up to 38 in Up to 43 in Up to 43 in Up to 60 in Up to 60 in Up to 60 in	(4.4 + 2.6L) μin (6.6 + 2.6L) μin (38 + 2.2L) μin (72 + 2L) μin (391 + 6.2L) μin (700 + 4.6L) μin	Comparison to Grade 0 Gage Blocks
Height Master ¹	Up to 12 in	65 μin	Comparison to Grade 0 Gage Blocks, Digital Indicator

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Plain Plug Gage ^{1,3}	Up to 8 in	$(15 + 2D) \mu\text{in}$	Comparison to Grade 0 Gage Blocks, P&W Supermicrometer [®]
Plain Ring Gage ^{1,3}	(0.04 to 8) in	$(8.6 + 3.6D) \mu\text{in}$	Comparison to Master Ring Gages, P&W Lab Master [®]
Thread Plug ^{1,3} Pitch Diameter	(0.06 to 6) in	$(58 + 3.4D) \mu\text{in}$	Comparison to Grade 0 Gage Blocks, P&W Supermicrometer [®] , Thread Wires
Major Diameter	(0.06 to 6) in	$(15 + 2D) \mu\text{in}$	
Thread Ring ^{1,3} Pitch Diameter	Up to 3 in	$(58 + 3.4D) \mu\text{in}$	Comparison to Thread Setting Plug
Minor Diameter	(0.046 to 0.2) in (0.2 to 3) in	$(15 + 2D) \mu\text{in}$ $(50 + 7.2D) \mu\text{in}$	Pin Gages, Bore Micrometer
Tapered Pipe Thread Plug ^{1,3}	Up to 3 in	$(130 + 1.5D) \mu\text{in}$	Comparison to Thread Wires, Sine Block, P & W Supermicrometer [®] , Height Master, Indicator
Tapered Sine Block ¹	(0.0625 to 3) in	71 μin	Comparison to 3/8 Gage Ball, P&W Supermicrometer [®] , Gage Blocks
Thread Measuring Wire ¹	(0.01 to 0.25) in (4 to 100) TPI	15.5 μin	Comparison to Class X Cylindrical Roll, P&W Supermicrometer [®]
Micrometer Standard ^{1,3}	Up to 6 in (6 to 20) in	$(35 + 1.9L) \mu\text{in}$ $(35 + 2.4L) \mu\text{in}$	Grade 0 Gage Blocks, P&W Supermicrometer [®]
Calipers ^{1,3} (5 μin resolution) (0.001) in resolution	Up to 60 in Up to 60 in	$(360 + 3.5L) \mu\text{in}$ $(550 + 16.7L) \mu\text{in}$	Comparison to Grade 0 Gage Blocks; Cylindrical Ring Gages
Test Indicator ¹ (50 μin resolution) (100 μin resolution) (500 μin resolution) (0.001 in resolution)	Up to 0.06 in Up to 0.06 in Up to 0.06 in Up to 0.06 in	51 μin 95 μin 470 μin 730 μin	Comparison to Grade 0 Gage Blocks

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Plunger Indicator ^{1,3} (20 μin resolution) (50 μin resolution) (100 μin resolution) (0.001 in resolution)	Up to 2 in Up to 2 in Up to 2 in Up to 2 in	(18 + 4.4L) μin (44 + 6.4L) μin (72 + 2.2L) μin (700 + 22L) μin	Comparison to Grade 0 Gage Blocks
Surface Plates ¹ Grades AA, A and B Overall Flatness	Up to 159 inDL	490 μin	In accordance with ASME B89.3.7 using Precision Level System
Local Area Flatness (repeat reading)	Up to 0.001 6 in	18 μin	Repeat-O-Meter
Squares, Steps, Angle Plates ¹	Up to 18 in	170 μin	Comparison to Squareness Checker
Spheres ¹	Up to 1 in	36 μin	Comparison to Grade 0 Gage Blocks, P&W Supermicrometer [®]
Rulers	Up to 6 in	70 μin	Comparison to OGP Smart Scope
Steel Rules	Up to 48 in	250 μin	Comparison to Optical Comparator
Thickness / Feeler Gages	(0.001 to 0.2) in	14 μin	Comparison to Grade 0 Gage Blocks, P&W Supermicrometer [®]
Optical Comparators ¹ Linearity	Up to 12 in	(90 + 7.6L) μin	Comparisons to Glass Scale
Squareness	Up to 6 in	190 μin	Square
Magnification	10X, 20X, 31.25X, 50X, 100 X	76 μin	Magnification Checker
Protractors ³ (Digital and Mechanical)	Up to 90°	7.5"	Comparison to Angle Blocks, Surface Plate
Angle Measure ³	Up to 90°	11"	Comparison to Gage Blocks, Surface Plate, Mu-Checker
Line Scales (Tape Measure and Rulers)	Up to 30 ft	0.036 in	Comparison to Laser Distance Meter

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Radius Gages (Leaf Type)	Up to 2 in	200 μ in	Direct Measure using Optical Comparator.

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure – Measure ¹	Up to 1 000 psig	3.2 psi	Comparison to Heise Gauge
Pressure – Source ¹	Up to 200 psig Up to 15 000 psig	0.25 psi 31 psi	Comparison to Pressure Calibrator, Deadweight Tester
Vacuum – Source ¹	(-30 to 0) inHg	0.017 inHg	Comparison to Manometer
Scales and Balances ¹	Up to 20 g	87 μ g	ASTM E617 Class 1 Weights and internal calibration procedure utilized in the calibration of the weighing system.
Scales ¹	Up to 1 000 lb	1.3 lb	NIST Class F Weights and internal calibration procedure utilized in the calibration of the weighing system.
Durometers Spring Force Only Type A	Up to 8.05 N	0.06 N	Partial Verification per ASTM D2240 using Durocalibrator
Type D	Up to 44.45 N	0.27 N	
Rockwell/Rockwell Superficial Hardness Testers ^{1,4,5}	62.1 HRA 71.1 HRA 81.8 HRA 46.5 HRC 45.4 HRC 65.9 HRC	0.31 HRA 0.2 HRA 0.28 HRA 0.45 HRC 0.37 HRC 0.32 HRC	Indirect Verification per ASTM E18 using Hardness Test Blocks.

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Rockwell/Rockwell Superficial Hardness Testers ^{1,4,5}	46.5 HRBw	0.37 HRBw	Indirect Verification per ASTM E18 using Hardness Test Blocks.
	62.2 HRBw	0.31 HRBw	
	95.5 HRBw	0.45 HRBw	
	90.2 HREw	0.33 HREw	
	89.7 HRFw	0.59 HRFw	
	96.4 HRFw	0.5 HRFw	
	101 HRFw	0.48 HRFw	
	88.1 HRHw	0.42 HRHw	
	70.1 HR15N	0.47 HR15N	
	83.2 HR15N	0.43 HR15N	
	91.8 HR15N	0.51 HR15N	
	47.6 HR30N	0.34 HR30N	
	64.5 HR30N	0.5 HR30N	
	80.4 HR30N	0.53 HR30N	
	72.8 HR15Tw	0.29 HR15Tw	
80.9 HR15Tw	0.32 HR15Tw		
91.8 HR15Tw	0.49 HR15Tw		
44.3 HR30Tw	0.6 HR30Tw		
62.4 HR30Tw	0.32 HR30Tw		
81.5 HR30Tw	0.31 HR30Tw		
Vickers Hardness Testers ^{1,5}	HV 0.5		Indirect Verification per ASTM E384 using Hardness Test Blocks.
	110 HV	3.4 HV	
	460 HV	9.6 HV	
	710 HV	13.7 HV	
	HV 1		
	110 HV	4.2 HV	
460 HV	7.9 HV		
710 HV	12 HV		
Brinell Hardness Testers ^{1,5}	HBW 2.5 / 187.5		Indirect Verification per ASTM E10 using Hardness Test Blocks.
	260 HBW	26 HBW	
	460 HBW	27 HBW	

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Brinell Hardness Testers ^{1,5}	HBW 5 / 750		Indirect Verification per ASTM E10 using Hardness Test Blocks.
	260 HBW	6.9 HBW	
	460 HBW	12.4 HBW	
	HBW 10 / 3 000		
	190 HBW	2.6 HBW	
Knoop Hardness Testers ^{1,5}	300 HBW	4.5 HBW	Indirect Verification per ASTM E384 using Hardness Test Blocks.
	430 HBW	8.2 HBW	
	HK 0.1		
	130 HK	5.4 HK	
	520 HK	17.2 HK	
	770 HK	21.3 HK	
	HK 0.5		
	110 HK	5 HK	
	460 HK	13.1 HK	
680 HK	16.1 HK		
HK 1	110 HK	4.5 HK	
	470 HK	15 HK	
	680 HK	18.1 HK	
Torque Tools	(25 to 250) lbf in	0.68 % of reading	Comparison to Norbar Torque Calibrator
	(25 to 250) lbf ft	0.67 % of reading	
	(100 to 1 000) lbf ft	0.68 % of reading	
Force – Compression and Tension ¹	(> 0 to 10) lbf	0.12 % of reading + 0.25 lbf	Comparison to NIST Class F Weights
	(10 to 100) lbf	0.12 % of reading + 0.5 lbf	
Force – Compression and Tension ¹	(100 to 1 124) lbf	0.06 % of reading	Comparison to Load Cells
	(1 000 to 3 000) lbf	0.05 % of reading	
	(3 000 to 11 240) lbf	0.05 % of reading	
	(10 000 to 20 000) lbf	0.002 % of reading	
	(20 000 to 50 000) lbf	0.3 % of reading	
Force – Compression (Only)	(50 000 to 300 000) lbf	0.1 % of reading	Comparison to Load Cells

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Measure ¹	(-35 to 200) °C	0.08 °C	Comparison to Reference Thermometer with PRT

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Humidity – Measure ¹	Up to 90 %RH	1.7 %RH	Comparison to Temperature/Humidity Indicator

Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Measure ¹	(1.1 to 10) Hz	9.5 mHz	Comparison to 8.5 Digit Multimeter
	(10 to 100) Hz	0.58 Hz	
	100 Hz to 1.1 kHz	0.37 Hz	
	(1.1 to 10) kHz	58 mHz	
	(10 to 100) kHz	36 Hz	
	100 kHz to 1.1 MHz	0.37 kHz	
	(1.1 to 10) MHz	0.58 kHz	
(10 to 15) MHz	6 kHz		
Tachometers ³ (Non-Contact Type)	(30 to 500 000) rpm	2.5 rpm	Comparison to Nova-Strobe PBL Stroboscope
Stopwatches & Timers (Push Button)	Up to 24 hr	55 ms	Comparison to Bench Timer
Electric Timers (Pulse)	1 ms to 24 hr	4.5 μs	Comparison to Frequency Counter

TESTING

Mechanical

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Comments
Hardness – Rockwell and Superficial Scales	ASTM E18	Sintered and Wrought Products	Wilson Rockwell Hardness Tester, Rockwell Scales – A, B, C, E, F, G, & H Superficial Scales – T & N

Mechanical

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Comments
Hardness – Brinell	ASTM E10	Sintered and Wrought Products	Newage Dyna Brinell HB500 to HB3 000
Microhardness – Vickers	ASTM E92	Sintered and Wrought Products	Mitutoyo MVK, Heavy Load Vickers (500 to 5 000) g
Microhardness – Knoop and Vickers	ASTM E384 ASTM B933	Sintered and Wrought Products	Buehler Micro Hardness Tester, Knoop Scale – (10 to 1 000) g Range Vickers Scale – (10 to 1 000) g Range

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. CMC for Electrical – DC/Low Frequency Calibrations do not include possible contributions to uncertainty from a “best available” unit under test.
3. L = length in inches; D = diameter in inches; DL = diagonal length in inches; rpm = revolutions per minute; " = arc-second.
4. The following statement is from the NIST recommended practice guide (Special Publication 960-5, page 60, section 8.3.3, paragraph 2) “Currently, there are no generally agreed upon U.S. or international methods for calculating the measurement uncertainty of a Rockwell hardness machine or the uncertainty in the certified value of standardized test blocks.”
5. The value shown in the Range column is a Nominal value. The Actual value will be reported at the time of calibration along with the associated Measurement Uncertainty.
6. Unless otherwise specified in the far-right column above, the laboratory utilizes internally written calibration procedures in the process of calibrating the parameters listed in this document.



Jason Stine, Vice President