CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Quasytech Inc.
42 Green Valley Drive, Unit-4
Kitchener, ON N2P 2C3 Canada

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

DIMENSIONAL MEASUREMENT

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 23 October 2025

Certificate Number: AD-3250

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Quasytech Inc.
42 Green Valley Drive, Unit-4
Kitchener, ON N2P 2C3 Canada
Nikharv Shah
905-962-9830

DIMENSIONAL MEASUREMENT

Valid to: **October 23, 2025** Certificate Number: **AD-3250**

1 Dimensional

Parameter	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional Measurement 1D	Up to 2 m	30.76 µm	Optical Portable CMM, HandyProbe
	Up to 2 m	61.22 µm	Optical Portable 3D scanner, Metrascan
	(0 to 25.4) mm	5.86 µm	Micrometer
	(0 to 304.8) mm	6.22 µm	Digital Caliper

2 Dimensional

Parameter	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional Measurement 2D	Up to 2 m	27.93 µm	Optical Portable CMM, HandyProbe
	Up to 2 m	36.66 µm	Optical Portable 3D scanner Metrascan

3 Dimensional

Parameter	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional Measurement 3D	Up to 2 m	27.93 µm	Optical Portable CMM, HandyProbe
	Up to 2 m	36.66 µm	Optical Portable 3D scanner Metrascan

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (k=2), corresponding to a confidence level of approximately 95%.

Notes:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. AD-3250.

Jason Stine, Vice President