CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Saturn Scale Systems Inc.
5350 Mainway, Units A3 & A4
Burlington, Ontario L7L 6A4

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 27 April 2027

Certificate Number: L2205

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Saturn Scale Systems, Inc.

5350 Mainway, Units A3 & A4
Burlington, ON L7L 6A4

David Marno, President 800-464-5983

CALIBRATION

ISO/IEC 17025 Accreditation Granted: **27 April 2025**

Certificate Number: **L2205** Certificate Expiry Date: **27 April 2027**

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Lab Balance [1] Class I and II (0.000 1 g resolution) (0.001 g Resolution)	 (0 to 110) g (0 to 120) g	 0.36 mg 0.69 mg	Comparison to ASTM E617 Class 1 Weights
Lab Balance [1] Class I or II (0.01 g Resolution)	(0 to 5 000) g	17 mg	Comparison to ASTM E617 Class 0 and 1 Weights
Commercial / Industrial Scales [1,2] (0.2 g resolution) (1 g resolution) (10 g resolution) (20 g resolution) (100 g resolution) (0.5 kg resolution) (1 kg resolution) (5 kg resolution) (10 kg resolution)	 (0 to 6) kg (0 to 30) kg (0 to 220) kg (0 to 500) kg (0 to 1 000) kg (0 to 2 000) kg (0 to 4 000) kg (0 to 50 000) kg (0 to 100 000) kg	 0.32 g 1.4 g 6.9 g 32 g 71 g 640 g 710 g 7.1 kg 14 kg	Comparison to Measurement Canada Government Certified Test Weights Calibrated in Accordance with Part III of the Weights and Measures Regulations (OIML Class M1)

1899 L Street NW, Suite 1100-A, Washington, DC 20036
414-501-5494
www.anab.org

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (k=2), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. Commercial / Industrial Scales including Platform Scales, Bench Scales, Floor Scales, Heavy Capacity Coil Scales, Counting Scales, Crane / Hanging Scales, Hopper / Tank Scales, Truck Scales, Off Highway Scales Vehicle Scales, Rail Car Scales or any scale with a platform base.

Jason Stine, Vice President

1899 L Street NW, Suite 1100-A, Washington, DC 20036
414-501-5494
www.anab.org